

# 检测报告

## TEST REPORT

报告编号:  
Report No.

J201511192555-02-0001

第 1 页 共 3 页  
Page of

委托方  
Client

HK LEE HING INDUSTRY CO., LIMITED

委托方地址  
Address

1F Junfeng Building, Gongle, Xixiang, Baoan District,  
Shenzhen, Guangdong, China

仪器名称  
Description

Spring Impact Hammer

型号/规格  
Model/Type

LX-IK07

制造厂  
Manufacturer

HK LEE HING INDUSTRY CO., LIMITED

出厂编号  
Serial No.

LX1503168099

管理号  
Asset No.

-----

检测日期

2015年03月14日

Date of Test

Y M D

样品接收日期

2015年03月13日

Date of Receipt

Y M D

批准人:

Approved Signatory

任平 (副主任)

核 验:

Inspected by

任平

检 测:

Tested by

潘广存



本实验室地址: 广东省广州市黄埔大道西平云路163号

3112

Address: No. 163. Pingyun Rd, West of HuangPu Ave. Guangzhou. Guangdong. China

联系电话: 020-38699959, 38699960, 38698685

Tel.: 020-38699959, 38699960, 38698685

传真: 020-38698685, 邮政编码: 510656

Fax: 020-38698685, Post Code: 510656

主页地址: [http:// www.grgtest.com](http://www.grgtest.com)

Web: <http:// www.grgtest.com>

电子邮件: [webmaster@grgtest.com](mailto:webmaster@grgtest.com)

E-mail: [webmaster@grgtest.com](mailto:webmaster@grgtest.com)

## 检 测 说 明

### DIRECTIONS OF TEST

报告编号: LX2015035502991

第 2 页 共 3 页  
Page of

**Report No.**

1. 本站出具的数据均可溯源至国家计量基准和国际单位制(SI)。  
 ( All data issued by this station are traced to National Primary Standards and International System of Units(SI).)
2. 本结果只与受检测样品有关(The results relate only to the tested).
3. 未经本实验室书面批准, 不得部分复制检测/试验报告。  
 (This test shall not be reproduced except in full , without the written approval of our laboratory.)
4. 委托单位如果对本次检测/试验有异议, 应于收到报告之日起十五天内向本实验室提出。  
 (If there is any objection concerning the test, the client should inform the calibration laboratory within 15 days from the date of the device under test report return to the client.)
5. 证书内页中“P”代表“Pass”, “F”代表“Fail”, “N/A”代表“Not Applicable”。  
 (In the datesheet."P" represents "Pass" and "F" represents "Fail" and "N/A" represents "Not Applicable".)
6. 证书内页中“\*\*\*\*E\*\*”代表“\*\*\*\*×10\*\*”, 例如“1.1E-6”代表“1.1×10<sup>-6</sup>”。  
 (In the datesheet."\*\*\*\*E\*\*" represents "\*\*\*\*×10\*\*", for example"1.1E-6" represents "1.1×10<sup>-6</sup>".)
7. 本次检测的技术依据文件及我司该项目CNAS认可范围, 超出范围的数据不在CNAS认可范围内:  
 (Reference documents for the Test and the CNAS accreditation range, Beyond the scope of the data is not within the accreditation by CNAS):  
 GRGJL.WI-LX-09-816-2008 《弹簧冲击器检测规范》

8. 本次检测使用的主要测量标准( Main Standards of Measurement Used in the Test.):

名称 / 型号 Description / Model	编 号 Serial No	证书编号 Certificate No	证书有效期 Due Date	技术特征 Technique Character
弹簧冲击器校准装置/CJ-XZ	CJ010207	LX2015021506646	2016-01-22	1.0%
半径样板/(1~25)mm	LXR01	CD2015010000109	2016-1-12	±(0.020-0.040)mm

9. 检测地点、环境条件 (Place and environmental conditions of the test):

地点 本实验室 Place	温度 22 °C Temperature	相对湿度 50 %RH Relative Humidity
------------------	-------------------------	----------------------------------

10. 建议下次检测日期(Due Date): 2016年07月08日

## 检 测 结 果

### RESULTS OF TEST

报告编号: LX2015035502991

第 3 页 共 3 页

Report No.

Page of

1. Appearance: Pass;

2. Impact Energy Test (J):

Nominal Value	Measured Average	Error	Tolerance	Conclusion (P/F)
2.0	1.980	+0.020	2.0±0.10	P

3. Impact Hammer Head Radius:

Specification (mm)	Measured Value (mm)	Conclusion (P/F)
10	10	P

(以下空白)  
(The below is blank)

备注:

Notes:

1 结论(Conclusion): 所检项目合格

2 测量结果的扩展不确定度: 冲击能量  $U_{rel} = 3.0\%/0.3\text{mm}$  ( $k=2$ );

The expanded uncertainty of the measurement results;

(依据《JJF1059.1-2012测量不确定度评定与表示》)

According to JJF1059.1-2012 Evaluation and Expression of Uncertainty in Measurement;